

Data sheet

12.10.2020

Semi-quantitative test strips QUANTOFIX Chlorine

Parameter*	Value*
REF	91317
Evaluable on reflectometer	No
Shelf life (from production)	2.5 Years
Scope of delivery	100 test strips in a tube, additional reagents, instruction leaflet in a box
Parameter	Chlorine
Gradation	0 • 1 • 3 • 10 • 30 • 100 mg/L Cl ₂
Method	Redox reaction
Color change	white → red-violet

*These data are typical values and must not be regarded as specification.

www.mn-net.com


MACHEREY-NAGEL GmbH & Co. KG · Neumann-Neander-Str. 6-8 · 52355 Düren · Germany

DE/international:

Tel.: +49 24 21 969-0

Fax: +49 24 21 969-199

E-mail: info@mn-net.com

CH:

Tel.: +41 62 388 55 00

Fax: +41 62 388 55 05

E-mail: sales-ch@mn-net.com

FR:

Tel.: +33 388 68 22 68

Fax: +33 388 51 76 88

E-mail: sales-fr@mn-net.com

US:

Tel.: +1 484 821 0984

Fax: +1 484 821 1272

E-mail: sales-us@mn-net.com